



Fourtec Declaration of Accuracy

Applicable for MicroLite: LITE5008 and LITE5016 Temperature Data Loggers

The MicroLite data logger is produced using Surface Mounted Device (SMD) assembling technique. Each product has been through complete quality assurance and calibration testing measures. All Fourtec production procedures are fully ISO 9001 certificated.

The LITE5008 and LITE5016 Temperature USB data logger is rated to the following temperature specifications.

Note: *The resolution information given below is averaged across the entire range.*

Temperature: Range -40 to 80 °C ; Accuracy ± 0.3 °C ; Resolution 0.1 °C

For further information please contact support@fourtec.com

Fourtec – Fourier Technologies.
Quality Assurance Department



Manufactured under ISO 9001:2008

www.fourtec.com