



fourtec Declaration of Accuracy

For MicroLite II External NTC Temperature Logger: LITE5032P-EXT

The MicroLite II data logger is produced using Surface Mounted Device (SMD) assembling technique. Each product has been through complete quality assurance and calibration testing measures. All Fourtec production procedures are fully ISO 9001 certificated.

The LITE5032P-EXT External Temperature USB data logger is rated to the following temperature specifications.

Note: *The resolution information given below is averaged across the entire range.*

Temperature:

Range: -50 to 150 °C;

Accuracy: From -50 to 60 °C ± 0.3 °C; From 60 to 150 °C ± 0.5 %

Resolution: 0.1 °C

For further information please contact support@fourtec.com

fourtec – Fourier Technologies.
Quality Assurance Department



Manufactured under ISO 9001:2000

www.fourtec.com