



fourtec Declaration of Accuracy

For MicroLite II External Voltage Logger: LITE5032P-V

The MicroLite II data logger is produced using Surface Mounted Device (SMD) assembling technique. Each product has been through complete quality assurance and calibration testing measures. All fourtec production procedures are fully ISO 9001 certificated.

The LITE5032P-V External Voltage USB data logger is rated to the following specifications.

Note: *The resolution information given below is averaged across the entire range.*

Current:

Range: 0 to 10 V

Accuracy: ± 5 mV

Resolution: 1 mV (logger LCD), 4 mV (software)

For further information please contact support@fourtec.com

fourtec – Fourier Technologies.
Quality Assurance Department



Manufactured under ISO 9001:2008

www.fourtec.com